

The ibidi labware is comprised of a variety of μ-Slides, μ-Dishes, and μ-Plates, which have all been designed for high-end microscopic analysis of fixed or living cells. The high optical quality of the ibidi Polymer Coverslip is similar to that of glass, enabling a variety of microscopy techniques with uncompromised resolution and choice of wavelength.

The μ-Slide 4 Well^{Ph+} (phase contrast plus) is a chambered coverslip with 4 wells. In contrast to the standard μ-Slide 4 Well, it provides a special intermediate plate for meniscus-free phase contrast and high-end fluorescence microscopy.

This document applies to the following product:

80446 **μ-Slide 4 Well^{Ph+} ibiTreat**

Material

The μ-Slide 4 Well^{Ph+} is made of a polymer that has the highest optical quality. The ibidi Polymer Coverslip bottom exhibits extremely low birefringence and autofluorescence, similar to that of glass. It is not possible to detach the bottom from the upper part. The slide is intended for one-time use and is not autoclavable, since it is only temperature-stable up to 80 °C/175 °F. Please note that gas exchange between the medium and the incubator's atmosphere occurs partially through the polymer coverslip, which should not be covered.

Optical Properties of Polymer Coverslip

Refractive index (589 nm)	1.52
Abbe number	56
Thickness	No. 1.5 (180 μm)
Material	Polymer




WARNING – The ibidi Polymer Coverslip is compatible with certain types of immersion oil only. A list of suitable oils can be found in the Section “Immersion Oil”.

Geometry

The μ-Slide 4 Well^{Ph+} provides a standard slide format according to ISO 8037/1.

Specifications

Outer dimensions (w × l)	25.5 × 75.5 mm ²
Number of wells	4
Dimensions of wells (w × l × h)	21.6 × 11.4 × 9.3 mm ³
Volume per well	700 μl
Liquid height	3.0 mm
Height with/without lid	10.8/9.5 mm
Growth area per well	2.5 cm ²
Coating area per well	5.9 cm ²
Bottom	ibidi Polymer Coverslip



Shipping and Storage

This product is sterilized and sealed in a gas-permeable packaging. The shelf life under proper storage conditions (in a dry place, no direct sunlight) is outlined in the following table.

Conditions	
Shipping conditions	Ambient
Storage conditions	RT (15–25 °C)

Shelf Life	
ibiTreat	36 months


Surface

The μ-Slide 4 Well^{Ph+} is available with an ibi-Treat surface.

The tissue culture-treated, hydrophilic ibiTreat surface of the ibidi Polymer Coverslip is ideal for culturing adherent cells. It ensures excellent cell adhesion without the necessity for any additional coatings. Nonetheless, extracellular matrix protein coatings can be applied to the ibiTreat surface without any restrictions, if required.

Filling and Handling

The μ-Slide 4 Well^{Ph+} suppresses the meniscus that interferes with the phase contrast effect in standard open wells. Openings near the corners allow for easy access to the wells for filling and aspirating liquids. To fill the wells, use a standard pipet and inject the cell suspension directly into one of the openings. Medium exchange can be easily performed by aspirating the entire volume and refilling with 700 μl per well.



TIP – The day before seeding the cells, we recommend placing both the cell medium and the μ-Slide in the incubator for equilibration. This helps prevent air bubble formation in the wells during incubation.

Coating

Detailed information about coatings is provided in [Application Note 08: Coating Protocols for ibidi Labware](#).

In short, specific coatings are possible following this protocol:

1. Prepare your coating solution according to the manufacturer's specifications. Adjust the concentration to a coating area of 5.9 cm² and a volume of 700 μl per well.
2. Apply 700 μl per well and leave it at room temperature for at least 30 minutes.
3. Aspirate the solution and wash with the recommended protein dilution buffer.
4. The coated slide is ready to be used. Be aware that allowing the coated surface to dry out is not recommended, as some coating proteins may degrade upon drying.

Seeding Cells

1. Trypsinize and count the cells as usual. Dilute the cell suspension to the desired concentration. Depending on your cell type, application of a 5–11 × 10⁴ cells/ml suspension should result in a confluent layer within 2–3 days.
2. Apply 700 μl cell suspension per well. Avoid shaking, as this will result in inhomogeneous cell distribution.
3. Cover the slide with the supplied lid. Incubate as usual (e.g., at 37 °C and 5% CO₂).

Insensitive cells can be left in their seeding medium for several days and grow to confluence there. However, optimal results might be achieved when the medium is changed every 2–3 days. For this, carefully aspirate the old medium and replace it by 700 μl fresh medium per well.

Microscopy

To image your cells, no special preparations are necessary. Living or fixed cells can be directly observed, preferably on an inverted microscope. The bottom cannot be removed. For optimal results in fluorescence microscopy and for storage of fixed and stained samples, ibidi provides mounting media that are optimized for ibidi labware:

Cat. No. 50001: [ibidi Mounting Medium](#)

Cat. No. 50011: [ibidi Mounting Medium with DAPI](#)

Chemical Compatibility

The following table provides some basic information on the chemical and solvent compatibility of the μ-Slide 4 Well^{Ph+}. For a full list of compatible solvents and more information on chemical compatibility, visit [ibidi.com/chemicals](https://www.ibidi.com/chemicals).

Chemical / Solvent	Compatibility
Methanol	Yes
Ethanol	Yes
Formaldehyde	Yes
Acetone	Yes, without lid
Mineral oil	No
Silicone oil	Yes
Immersion oil	See Section “Immersion Oil”

Immersion Oil



WARNING – When using oil immersion objectives with the ibidi Polymer Coverslip, use only the immersion oils specified in the table below. The use of any non-recommended oil could damage the ibidi Polymer Coverslip. The resulting leakage may harm objectives and microscope components. All immersion oils that are not listed in the table below should be considered as non-compatible.

Company	Product	Ordering No.	Lot Number	Test Date
ibidi	ibidi Immersion Oil 2	50102	24-07-04	07/2024
Cargille	Type A	16482	100592	01/2017
Cargille	Type HF	16245	92192	01/2017
Carl Roth	Immersion oil	X899.1	414220338	01/2017
Leica	Immersion Liquid	11513859	n.a.	03/2023
Leica	Immersion Liquid Type G	11513910	n.a.	04/2024
Nikon	Immersion Oil F2 30cc	MXA22192	n.a.	01/2020
Nikon	Silicone Immersion Oil 30cc	MXA22179	20191101	01/2020
Olympus	Silicone Immersion Oil	SIL300CS-30CC	N4190800	01/2017
Zeiss	Immersol 518 F	444960-0000	220211	03/2023
Zeiss	Immersol 518 F (30 °C)	444970-9010	220816	03/2023
Zeiss	Immersol 518 F (37 °C)	444970-9000	220302	03/2023
Zeiss	Immersol W 2010	444969-0000	101122	04/2012
Zeiss	Immersol Sil 406	444971-9000	80730	03/2023
Zeiss	Immersol G	462959-9901	211117	03/2023

For research use only!

Further information can be found at ibidi.com. For questions and suggestions, please contact us by e-mail at info@ibidi.com or by telephone at +49 (0)89/520 4617 0.
© ibidi GmbH, Lochhamer Schlag 11, 82166 Gräfelfing, Germany.