



Product Service

Attestation of Conformity

No. N8 13 09 50336 006

Holder of Certificate: **ibidi GmbH**Am Klopferspitz 19
82152 Martinsried
GERMANY**Product:** **Laboratory equipment****Model(s):** **ibi pump 2****Parameters:**Rated voltage: 14 VDC
Rated current: 1 A
Protection class: IIIPower supply unit: SPU26-106
(Ratings: 100-240V, 47-63 Hz, 0.7A, LPS)**Tested according to:** EN 61010-1:2010

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2006/95/EC relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.: 028-713020174-100**Date,** 2013-09-19
(Erich Thurner)

CE After preparation of the necessary technical documentation as well as the EC conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.

Page 1 of 1